

## New England Biolabs Certificate of Analysis

**Product Name:** NEBNext® rRNA Depletion Kit (Human/Mouse/Rat)  
**Catalog Number:** E6310S  
**Packaging Lot Number:** 10223794  
**Expiration Date:** 01/2026  
**Storage Temperature:** -20°C  
**Specification Version:** PS-E6310S/L/X/B v1.0

NEBNext® rRNA Depletion Kit (Human/Mouse/Rat) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E6318-2VIAL	NEBNext® RNase H	10223803	Pass
E6317-2VIAL	Nuclease-free Water	10223801	Pass
E6316-2VIAL	DNase I (RNase-free)	10223800	Pass
E6315-2VIAL	DNase I Reaction Buffer	10223799	Pass
E6314-2VIAL	NEBNext® Probe Hybridization Buffer	10223798	Pass
E6313-2VIAL	NEBNext® rRNA Depletion Solution	10223796	Pass
E6312-2VIAL	RNase H Reaction Buffer	10223795	Pass

Assay Name/Specification	Lot # 10223794
<b>* Individual Product Component Note</b> Standard Quality Control Tests are performed for each component included in NEBNext® rRNA Depletion Kit (Human/Mouse/Rat) and meet the designated specifications.	<b>Pass</b>
<b>Functional Testing (rRNA Depletion)</b> The NEBNext rRNA Depletion Kit is functionally validated using commercially available human RNA (e.g. UHRR). After treatment with the kit, a high throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with less than 10% rRNA.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
04 Mar 2024

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
11 Apr 2024