

be INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	NEBNext® Globin <(>&<)> rRNA Depletion Kit (Human/Mouse/Rat)
Catalog Number:	E7750X
Lot Number:	10049674
Expiration Date:	05/2021
Storage Temperature:	-20°C
Specification Version:	PS-E7750S/L/X v1.0
Expiration Date: Storage Temperature:	05/2021 -20°C

NEBNext® Globin <(>&<)> rRNA Depletion Kit (Human/Mouse/Rat) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7753-4VIAL	NEBNext® DNase I	10042866	Pass
E7752-4VIAL	NEBNext® Thermostable RNase H	10042867	Pass
E7751-4VIAL	NEBNext® Globin & rRNA Depletion Solutio	10042868	Pass
E6317-4VIAL	Nuclease-free Water	10046670	Pass
E6315-4VIAL	DNase I Reaction Buffer	10042870	Pass
E6314-4VIAL	NEBNext® Probe Hybridization Buffer	10042872	Pass
E6312-4VIAL	RNase H Reaction Buffer	10042869	Pass

Assay Name/Specification	Lot # 10049674
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) and meet the designated specifications.	Pass
Functional Testing (Globin and rRNA Depletion) The NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) is functionally validated using commercially available human RNA extracted from blood. After treatment with the kit, a high throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with less than 10% rRNA and less than 5% globin RNA.	Pass

This product has been tested and shown to be in compliance with all specifications.





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

notin ~

Christine Sumner Production Scientist 25 Jul 2019

Michae on. V

Michael Tonello Packaging Quality Control Inspector 25 Jul 2019

