

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: T4 RNA Ligase 2

Catalog Number: M0239L Concentration: 10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to ligate 0.4

μg of an equimolar mix of a 23-mer and 17-mer RNAs in a total

reaction volume of 20 µl in 30 minutes at 37°C.

Lot Number: 10029433
Expiration Date: 07/2020
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl, 50 mM KCl, 0.1 mM DTT, 0.1 mM EDTA, 35 mM (NH4)2SO4,

50 % Glycerol, (pH 7.5 @ 25°C)

Specification Version: PS-M0239S/L v1.0

T4 RNA Ligase 2 Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
M0239LVIAL	T4 RNA Ligase 2 (dsRNA Ligase)	10015139	Pass	
B0239SVIAL	T4 RNA Ligase 2 Reaction Buffer	10028358	Pass	

Assay Name/Specification	Lot # 10029433
Endonuclease Activity (Nicking) A 50 μl reaction in T4 RNA Ligase Reaction Buffer containing 1 μg of supercoiled PhiX174 DNA and a minimum of 100 units of T4 RNA Ligase 2 (dsRNA Ligase) incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 μl reaction in T4 RNA Ligase Reaction Buffer containing 1 μg of a mixture of single and double-stranded [ ³H] E. coli DNA and a minimum of 100 units of T4 RNA Ligase 2 (dsRNA Ligase) incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Phosphatase Activity (pNPP) A 200 µl reaction in 1M Diethanolamine, pH 9.8, 0.5 mM MgCl2 containing 2.5 mM p-Nitrophenyl Phosphate (pNPP) and a minimum of 100 units of T4 RNA Ligase 2 (dsRNA Ligase) incubated for 4 hours at 37°C yields <0.00001 unit of alkaline phosphatase activity as determined by spectrophotometric analysis.	Pass



M0239L / Lot: 10029433

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Bo Wu

Production Scientist

18 Jul 2018

Michael Tonello

Packaging Quality Control Inspector

20 Nov 2018

