

240 County Road Ipswich, MA 01938-2723

Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: RNase Inhibitor, Murine

Catalog Number: M0314S
Concentration: 40,000 U/ml

Unit Definition: One unit is defined as the amount of Murine RNase Inhibitor required

to inhibit the activity of 5ng of RNase A by 50%. Activity is measured by the inhibition of hydrolysis of cytidine 2', 3'-cyclic

monophosphate by RNase A.

Lot Number: 10037465
Expiration Date: 03/2021
Storage Temperature: -20°C

Storage Conditions: 50 mM KCI, 20 mM HEPES (pH 7.6), 8 mM DTT, 50 % Glycerol

Specification Version: PS-M0314S/L v1.0

RNase Inhibitor, Murine Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
M0314SVIAL	RNase Inhibitor, Murine	10037466	Pass	

Assay Name/Specification	Lot # 10037465
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA	Pass
and a minimum of 40 units of RNase Inhibitor, Murine is incubated at 37°C. After	
incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	
Endonuclease Activity (Nicking)	Pass
A 50 µl reaction in NEBuffer 4 containing 1 µg of supercoiled PhiX174 DNA and a minimum of 40 units of RNase Inhibitor, Murine incubated for 4 hours at 37°C results	
in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	
Exonuclease Activity (Radioactivity Release)	Pass
A 50 µl reaction in NEBuffer 4 containing 1 µg of a mixture of single and	
double-stranded [3H] E. coli DNA and a minimum of 200 units of RNase Inhibitor,	
Murine incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	
Latent RNase Activity (Extended Digest)	Pass
A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA	
and a minimum of 40 units of heat inactivated RNase Inhibitor, Murine is incubated	



M0314S / Lot: 10037465

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Dongxian Yue Production Scientist

12 Mar 2019

Michael Tonello

Packaging Quality Control Inspector

14 Mar 2019