

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: T4 RNA Ligase 2, truncated KQ

Catalog Number: M0373S
Concentration: 200,000 U/ml

Unit Definition: 200 units is defined as the amount of enzyme required to give 80%

ligation of a 31-mer RNA to the pre-adenylated end of a 17-mer DNA

in a total reaction volume of 20 µl in 1 hour at 25°C.

Packaging Lot Number: 10052062
Expiration Date: 08/2021
Storage Temperature: -20°C

Storage Conditions: 100 mM NaCl , 10 mM Tris-HCl , 1 mM DTT , 0.1 mM EDTA , 50 %

Glycerol, (pH 7.4 @ 25°C)

Specification Version: PS-M0373S/L v2.0

T4 RNA Ligase 2, truncated KQ Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
M0373SVIAL	T4 RNA Ligase 2, truncated KQ	10052061	Pass	
B1004SVIAL	PEG 8000	10045988	Pass	
B0216SVIAL	T4 RNA Ligase Reaction Buffer	10048447	Pass	

Assay Name/Specification	Lot # 10052062
Endonuclease Activity (Nicking) A 50 μl reaction in T4 RNA Ligase Reaction Buffer containing 1 μg of supercoiled PhiX174 DNA and a minimum of 200 units of T4 RNA Ligase 2, truncated KQ incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 μl reaction in T4 RNA Ligase Reaction Buffer containing 1 μg of a mixture of single and double-stranded [ ³H] E. coli DNA and a minimum of 200 units of T4 RNA Ligase 2, truncated KQ incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Phosphatase Activity (pNPP) A 200 µl reaction in 1M Diethanolamine, pH 9.8, 0.5 mM MgCl2 containing 2.5 mM p-Nitrophenyl Phosphate (pNPP) and a minimum of 200 units of T4 RNA Ligase 2, truncated KQ incubated for 4 hours at 37°C yields <0.00001 unit of alkaline phosphatase activity as determined by spectrophotometric analysis.	Pass



M0373S / Lot: 10052062

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Bo Wu

**Production Scientist** 

24 Jun 2019

Michael Tonello

Packaging Quality Control Inspector

25 Nov 2019

