

## New England Biolabs Certificate of Analysis

**Product Name:** Acul  
**Catalog Number:** R0641S  
**Concentration:** 5,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.  
**Lot Number:** 10026462  
**Expiration Date:** 10/2020  
**Storage Temperature:** -20°C  
**Storage Conditions:** 100 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA  
**Specification Version:** PS-R0641S/L v2.0

Acul Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0641SVIAL	Acul	10026463	Pass
B9003SVIAL	S-adenosylmethionine (SAM)	10033230	Pass
B7204SVIAL	CutSmart® Buffer	10031565	Pass

Assay Name/Specification	Lot # 10026462
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 5 Units of Acul incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
<b>Protein Purity Assay (SDS-PAGE)</b> Acul is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass
<b>Endonuclease Activity (Nicking)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 5 units of Acul incubated for 4 hours at 37°C results in <50% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 25 units of Acul incubated for 4 hours at 37°C releases <0.2% of the total radioactivity.	Pass

Assay Name/Specification	Lot # 10026462
<p><b>Ligation and Recutting (Terminal Integrity)</b> After a 10-fold over-digestion of Lambda DNA with Acul, ~50% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, ~75% can be recut with Acul.</p>	<p><b>Pass</b></p>

This product has been tested and shown to be in compliance with all specifications.



Stephanie Cornelio  
Production Scientist  
25 Oct 2018



Michael Tonello  
Packaging Quality Control Inspector  
16 Jan 2019