

## New England Biolabs Certificate of Analysis

**Product Name:** Monarch® HMW DNA Extraction Kit for Tissue  
**Catalog Number:** T3060L  
**Packaging Lot Number:** 10152960  
**Expiration Date:** 06/2024  
**Storage Temperature:** 25°C  
**Specification Version:** PS-T3060S/L v2.0


Monarch® HMW DNA Extraction Kit for Tissue Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T3062-1	Monarch® Protein Separation Solution	10152825	Pass
T3061-1	Monarch® HMW gDNA Tissue Lysis Buffer	10152822	Pass
T3056-1	Monarch® gDNA Elution Buffer II	10152815	Pass
T3018-2	Monarch® RNase A	10111844	Pass
T3015-1	Monarch® gDNA Wash Buffer	10153311	Pass
T3005-1	Monarch® DNA Capture Beads (100)	10152785	Pass
T3004-1	Monarch® Bead Retainers (50)	10142460	Pass
T3003-1	Monarch® 2 ml Tubes (50)	10152771	Pass
T3002-1	Monarch® Pestles (50)	10152769	Pass
T3001-1	Monarch® Pestle Tubes (50)	10152756	Pass
T2018-1	Monarch® Collection Tubes II	10079601	Pass
P8107SVIAL	Proteinase K, Molecular Biology Grade	10152683	Pass

Assay Name/Specification	Lot # 10152960
<b>Functional Testing (Genomic DNA Yield and Integrity Analysis) (Tissue)</b> Genomic DNA is purified from 8 samples of NEB 10-beta Competent E. coli using the High Molecular Weight DNA Extraction from Bacteria protocol. Each set of reagents is functionally tested to ensure successful isolation of gDNA from 1 x 10e9 cells. Yield of genomic DNA is ≥6 µg in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples, and DIN values are ≥9 in ≥80% of the samples. In ≥ 80% of samples, ≥ 75% of the DNA is 50 kb in size or greater.	<b>Pass</b>
<b>* Individual Product Component Note</b> Standard Quality Control Tests are performed for each component included in Monarch® HMW DNA Extraction Kit for Tissue and meet the designated specifications.	<b>Pass</b>


This product has been tested and shown to be in compliance with all specifications.



One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Ana Egana  
Production Scientist  
21 Sep 2022



Michael Burzyk  
Packaging Quality Control Inspector  
21 Sep 2022